

All data taken at the Pacific Northwest National Laboratory

FTS Operators: Schuyler P. Lockwood

Data Analysis: Tracy J. Baker

**Composite spectrum for:** Tropane

- First Column: Position in wavenumber ( $\text{cm}^{-1}$ )
- Second column: Real refractive index  $n(\tilde{\nu})$  (dispersion index)
- Third column: Imaginary refractive index,  $k(\tilde{\nu})$  (absorption index per unit length in centimeters)

Where the complex refractive index  $\hat{n} = n(\tilde{\nu}) + ik(\tilde{\nu})$

Following Bertie (in the references below) we define the absorbance as  $A = -\log_{10}(I/I_0)$  and the linear absorption coefficient  $K = A/d$ , where  $d$  is the path length. The connection between the imaginary refractive index and the absorbance coefficient arises from the following:  $2.303K = 4\pi\tilde{\nu}k$

See the following references for a detailed description of terms and units:

- 1) Bertie, J. E., Zhang, S. L., Eysel, H. H., Baluja, S., & Ahmed, M. K. (1993). Infrared Intensities of Liquids XI: Infrared Refractive Indices from 8000 to  $2\text{ cm}^{-1}$ , Absolute Integrated Intensities, and Dipole Moment Derivatives of Methanol at  $25^\circ\text{C}$ . *Appl. Spec.*, 47(8), 1100-1114 doi:10.1366/0003702934067973
- 2) Bertie, J. E., Zhang, S. L., & Keefe, C. D. (1995). Measurement and use of absolute infrared absorption intensities of neat liquids. *Vibrational Spectroscopy*, 8(2), 215-229. doi:10.1016/0924-2031(94)00038-i

**Sample:**

- Chemical name, formula and CAS number: Tropane,  $\text{C}_8\text{H}_{15}\text{N}$ , [529-17-9]
- IUPAC name: 8-Methyl-8-azabicyclo[3.2.1]octane
- Synonyms: 2,3-Dihydro-8-methylnortropidine
- Physical properties: FW = 125.21 g/mole; mp = n/a; bp =  $167\text{--}169^\circ\text{C}$ ;  $\rho = 0.931\text{ g/cm}^3$
- Supplier and stated purity: Aldrich, 98% (Lot # MKBF3966V)
- Temperature of sample:  $24^\circ\text{C}$  ( $\pm 1^\circ\text{C}$ )
- Individual samples were measured at the following path lengths: MIR: 0.81, 3.73, 10.3, 12.5, 14.7, 26.4, 49.1, 107, 211 and 464 micrometers ( $\mu\text{m}$ ); NIR: 108, 211, 462, 925 and 2095  $\mu\text{m}$ . Final data are a composite of these spectra.
- Sample cell window material: MIR = potassium bromide (KBr) except potassium chloride (KCl) for the 107 and 211  $\mu\text{m}$  cells; NIR = KCl except KBr for the 462 and 925  $\mu\text{m}$  cells.
- Preparation: None

**NIR Instrument Parameters:**

- Bruker IFS66V, purged with UHP nitrogen
- Spectral range:  $12,000$  to  $3,000\text{ cm}^{-1}$  (0.83 to 3.33 microns)
- NIR source: Quartz tungsten bulb
- Beamsplitter: Quartz
- Detector: DLTGS at room temperature
- Aperture: 5 mm
- Folding limits:  $15798.0$  to  $0\text{ cm}^{-1}$

**MIR Instrument Parameters:**

- Tensor II with stage rotated 6 degrees, purged with UHP nitrogen
- Spectral range:  $7,800$  to  $400\text{ cm}^{-1}$  (1.282 to 25 microns)
- IR source: Silicon carbide glow bar
- Beamsplitter: Broadband potassium bromide (KBr)
- Detector: DTGS at room temperature
- Aperture: 3 mm
- Folding limits:  $11669.94$  to  $0\text{ cm}^{-1}$

### NIR/MIR Instrument Parameters:

- Instrument resolution:  $2.0\text{ cm}^{-1}$
- Number of interferograms averaged per single channel spectrum: 128
- Apodization: Norton-Beer, Medium
- Phase correction: Mertz
- Scanner velocity: 10 kHz; 7.5 kHz (MIR)
- Interferogram zerofill: 4x
- Spectral interval after zerofilling:  $0.4822\text{ cm}^{-1}$

a)



b)



Figure 1: The Bruker IFS66V FTIR (a) and Tensor II (b).

### Measured Refractive Index:

The refractive index for Tropane was measured at  $25\text{ }^{\circ}\text{C}$  using an Atago model DR-M2/1550 Abbe refractometer. Notch filters were employed in front of a white light source to make measurements at multiple wavelengths. An infrared viewer from Atago was used to detect signal at  $1550\text{ nm}$ . The temperature was controlled to match that in the sample compartment of the FTIR using a heated circulating bath.

|          |              |         |              |         |              |
|----------|--------------|---------|--------------|---------|--------------|
| 480 nm:  | $n = 1.4820$ | 486 nm: | $n = 1.4812$ | 546 nm: | $n = 1.4770$ |
| 589 nm:  | $n = 1.4751$ | 644 nm: | $n = 1.4729$ | 656 nm: | $n = 1.4728$ |
| 1550 nm: | $n = 1.4558$ |         |              |         |              |

The refractive index,  $n$ , vs. wavelength in microns,  $\lambda$ , was fit to an equation similar to that of Sellmeier:

$$n(\lambda) = \{a + b/(\lambda^2 - c)\}^{1/2}$$

The resulting best-fit equation was used to find the refractive index at the highest energy data points in our experimental spectra. For Tropane, the results were

$$\begin{aligned} n(7,800\text{ cm}^{-1}) &= 1.4581 \text{ at } 25\text{ }^{\circ}\text{C} \text{ for MIR data and} \\ n(12,000\text{ cm}^{-1}) &= 1.4661 \text{ at } 25\text{ }^{\circ}\text{C} \text{ for NIR and merged data.} \end{aligned}$$

### Post Processing and Related Parameters:

For the MIR, a composite spectrum was created from 10 absorbance spectra (base-10) taken at 10 path lengths: 0.81, 3.73, 10.3, 12.5, 14.7, 26.4, 49.1, 107, 211 and 464 micrometers ( $\mu\text{m}$ ). These data were collected with the sample stage rotated by  $-6^\circ$  to minimize artifacts resulting from back-reflection into spectrometer (see Johnson et al., *Appl. Spectrosc.*, 76(5) 620-624, 2021). For the NIR, a composite spectrum was created from 5 absorbance spectra (base-10) taken at 5 path lengths: 108, 211, 462, 925 and 2095  $\mu\text{m}$ . The same cells and liquid fills for the  $\sim 100$ , 200 and 500  $\mu\text{m}$  path lengths were used for both spectral ranges. At each path length several spectra were measured and the results averaged for better signal to noise. The measured cell lengths were adjusted using Beer's law plots in which the NIR and MIR data were analyzed independently.

- 1) The imaginary part of the refractive index, or  $k$  vector, was determined for each absorbance file as per Bertie's program "RNJ46A" (see reference above). This takes into account the reflective losses due to the KBr and/or KCl windows.
- 2) A composite  $k$  vector is created via a classical, weighted, linear, least squares fit using the output files of program "RNJ46A": Intercept=0, slope is fitted, individual absorbance values weighted by  $T^2$  (transmission squared), all absorbance values  $\geq 2.5$  are given zero weight. For the MIR, eight composite vectors were created and merged by hand.
  - a) The first  $k$  vector used the results from the 211 and 464  $\mu\text{m}$  cells. This  $k$  vector determined the final values for the range from 7800 to 3000  $\text{cm}^{-1}$ .
  - b) The second  $k$  vector used the results from 10.3 and 14.7 through 107  $\mu\text{m}$  cells. This  $k$  vector determined the final values for the range from 3000 to 2960  $\text{cm}^{-1}$ .
  - c) The third  $k$  vector used the results from the 10.3  $\mu\text{m}$  cell. This  $k$  vector determined the final values for the range from 2960 to 2900  $\text{cm}^{-1}$ .
  - d) The fourth  $k$  vector used the results from the 10.3 through 26.4  $\mu\text{m}$  cells. This  $k$  vector determined the final values for the range from 2900 to 2780  $\text{cm}^{-1}$ .
  - e) The fifth  $k$  vector used the results from the 49.1 and 107  $\mu\text{m}$  cells. This  $k$  vector determined the final values for the range from 2780 to 2750  $\text{cm}^{-1}$ .
  - f) The sixth  $k$  vector used the results from the 464  $\mu\text{m}$  cell. This  $k$  vector determined the final values for the range from 2750 to 1770  $\text{cm}^{-1}$  and 500 to 370  $\text{cm}^{-1}$ .
  - g) The seventh  $k$  vector used the results from the 107 and 211  $\mu\text{m}$  cells. This  $k$  vector determined the final values for the range from 1770 to 1500  $\text{cm}^{-1}$ .
  - h) The eighth  $k$  vector used the results from the 12.5 through 26.4  $\mu\text{m}$  cells. This  $k$  vector determined the final values for the range from 1500 to 500  $\text{cm}^{-1}$ .
- 3) A frequency correction was applied to the resulting composite MIR  $k$  vector.
  - a) Frequency correction (already applied):  $\tilde{\nu}(\text{corrected}) = [\tilde{\nu}(\text{instrument}) * 0.99991 + 0.047]$  as determined by comparing measured atmospheric spectral lines ( $\text{H}_2\text{O}$  and  $\text{CO}_2$ ) to values from the Northwest Infrared Spectral Library Database.
- 4) For the NIR, five composite vectors were created and merged by hand.
  - a) The first  $k$  vector used the results from the 2095  $\mu\text{m}$  cell. This  $k$  vector determined the final values for the range from 12,000 to 5900  $\text{cm}^{-1}$  and 5700 to 4500  $\text{cm}^{-1}$ .
  - b) The second  $k$  vector used the results from the 925 and 2095  $\mu\text{m}$  cells. This  $k$  vector determined the final values for the range from 5900 to 5700  $\text{cm}^{-1}$ .
  - c) The third  $k$  vector used the results from the 462 and 925  $\mu\text{m}$  cells. This  $k$  vector determined the final values for the range from 4500 to 3660  $\text{cm}^{-1}$ .
  - d) The fourth  $k$  vector used the results from the 211 and 462  $\mu\text{m}$  cells. This  $k$  vector determined the final values for the range from 3660 to 3000  $\text{cm}^{-1}$ .
  - e) The fifth  $k$  vector used the results from the 108 through 462  $\mu\text{m}$  cells. This  $k$  vector determined the final values for the range from 3000 to 400  $\text{cm}^{-1}$ .
- 5) The resulting composite NIR  $k$  vector and the refractive index at 10,000  $\text{cm}^{-1}$  were used to create the real or  $n$  vector using the Kramers-Kronig relation, as per Bertie's program "LZZKTB."
  - a) Frequency correction (already applied):  $\tilde{\nu}(\text{corrected}) = [\tilde{\nu}(\text{instrument}) * 1 + 0.082]$  as determined by comparing measured atmospheric spectral lines ( $\text{H}_2\text{O}$  and  $\text{CO}_2$ ) to values from the Northwest Infrared Spectral Library Database.
- 6) The MIR data were mapped onto the NIR x-axis using an interpolation routine, i.e. the Make Compatible

command in OPUS 5.5. Then the composite MIR and NIR  $k$  vectors were merged around  $3000\text{ cm}^{-1}$  to generate a final composite  $k$  vector across the entire spectral range. A weighted average, with the weight of the MIR vector increasing linearly from 0 to 100%, was used in the overlapping spectral region. The resulting composite  $k$  vector and the refractive index at  $12,000\text{ cm}^{-1}$  were used to create the final  $n$  vector using the Kramers-Kronig relation, as per Bertie's program "LZZKTB."

**Photograph of Sample Tropane:**



Figure 2: Tropane in Aldrich container for NIR and MIR measurements.